



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Accredited Calibration Services, Inc. (Marsh Metrology)
2-1016C Sutton Drive
Burlington ON L7L 6B8 Canada

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

AC-1172
Certificate Number

ANAB Approval

Certificate Valid: 04/03/2018-05/27/2020
Version No. 006 Issued: 04/03/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



ANSI-ASQ National Accreditation Board

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 AND ANSI/NCSL Z540-1-1994 (R2002)

Accredited Calibration Services, Inc. (Marsh Metrology)

2-1016C Sutton Drive
Burlington, ON L7L 6B8 Canada
Ron Bake 905-331-9783
Ron.bake@marshmetrology.com www.marshmetrology.com

CALIBRATION

Valid to: May 27, 2020

Certificate Number: AC-1172

Chemical Quantities

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
pH Meters	(4, 7 and 10) pH	0.012 pH	Standard Buffer Solutions

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Voltage - Source	Up to 330 mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V 330 V to 1 kV	14 μ V/V + 1.2 μ V 6.3 μ V/V + 10 μ V 7.7 μ V/V + 81 μ V 12 μ V/V + 0.78 mV 14 μ V/V + 1.4 mV	Multifunction Calibrator
DC Voltage - Measure	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V 100 V to 1 kV	9.6 μ V/V + 1.1 μ V 3.8 μ V/V + 10 μ V 8.4 μ V/V + 3.3 μ V 10 μ V/V + 38 μ V 10 μ V/V + 0.13 mV	High Resolution DMM
	Up to 6 kV (6 to 20) kV (20 to 35) kV Up to 150 kV	10 mV/V + 0.6 V 20 mV/V + 2.4 V 90 mV/V + 51 V 5.1 mV/V + 6.9 V	DMM with High Voltage Probe



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Thermocouple Simulation and Measure	J-type thermocouple (63 to 1 473) K (-210 to 1 200) °C K-type thermocouple (73 to 1 645) K (-200 to 1 372) °C S-type thermocouple (273 to 1 673) K (0 to 1 400) °C T-type thermocouple (23 to 673) K (-250 to 400) °C E-type thermocouple (23 to 1 273) K (-250 to 1 000) °C N-type thermocouple (73 to 1 573) K (-200 to 1 300) °C	0.24 K (0.24 °C) 0.25 K (0.25 °C) 0.52 K (0.52 °C) 0.25 K (0.25 °C) 0.43 K (0.43 °C) 0.37 K (0.37 °C)	Multifunction Calibrator
AC Voltage - Source	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz 330 mV to 3.3 V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.62 mV/V + 4.8 µV 0.12 mV/V + 4.7 µV 0.15 mV/V + 4.8 µV 0.78 mV/V + 4.7 µV 2.7 mV/V + 9.4 µV 6.2 mV/V + 39 µV 0.39 mV/V + 6.8 µV 0.11 mV/V + 7.1 µV 0.12 mV/V + 7.4 µV 0.27 mV/V + 6.7 µV 0.62 mV/V + 25 µV 1.6 mV/V + 54 µV 0.23 mV/V + 43 µV 0.11 mV/V + 66 µV 0.14 mV/V + 58 µV 0.23 mV/V + 42 µV 0.54 mV/V + 0.1 mV 1.9 mV/V + 0.47 mV	Multifunction Calibrator



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage - Source	(3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz 330 V to 1 kV 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.23 mV/V + 0.53 mV 0.11 mV/V + 0.53 mV 0.19 mV/V + 0.49 mV 0.27 mV/V + 0.5 mV 0.7 mV/V + 1.3 mV 0.11 mV/V + 2.4 mV 0.15 mV/V + 5.5 mV 0.19 mV/V + 5.4 mV 0.23 mV/V + 5.3 mV 1.6 mV/V + 39 mV 0.23 mV/V + 8.1 mV 0.19 mV/V + 8 mV 0.23 mV/V + 8.2 mV	Multifunction Calibrator
AC Voltage - Measure	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz 100 mV to 1 V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (1 to 10) V (1 to 40) Hz 40 H to 1 kHz (1 to 20) kHz (20 to 50) kHz	0.47 mV/V + 4 μ V 0.14 mV/V + 3 μ V 0.22 mV/V + 3 μ V 0.89 mV/V + 2.7 μ V 70 μ V/V + 4.3 μ V 70 μ V/V + 2.1 μ V 0.14 mV/V + 2.1 μ V 0.3 mV/V + 2.1 μ V 70 μ V/V + 40 μ V 70 μ V/V + 21 μ V 0.14 mV/V + 21 μ V 0.3 mV/V + 23 μ V 70 μ V/V + 0.4 mV 70 μ V/V + 0.22 mV 0.14 mV/V + 0.21 mV 0.3 mV/V + 0.21 mV	High Resolution DMM, Multifunction Calibrator



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage - Measure	(10 to 100) V (1 to 40) Hz 40 H to 1 kHz (1 to 20) kHz (20 to 50) kHz 100 V to 1 kV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz Up to 6 kV 60 Hz (6 to 35) kV 60 Hz Up to 150 kV 1 kHz	0.2 mV/V + 4 mV 0.2 mV/V + 2 mV 0.2 mV/V + 2.1 mV 0.35 mV/V + 2.1 mV 0.4 mV/V + 40 mV 0.4 mV/V + 20 mV 0.6 mV/V + 20 mV 1.1 mV/V + 79 mV 10 mV/V + 5 V 51 mV/V + 9 V 8.5 mV/V + 38 V	High Resolution DMM, Multifunction Calibrator
DC Current - Source	Up to 330 µA 330 µA to 3.3 mA (3.3 to 330) mA (33 to 330) mA 330 mA to 1.1 A (1.1 to 3) A (3 to 11) A	59 µA/A + 61 nA 73 µA/A + 57 nA 77 µA/A + 0.21 µA 75 µA/A + 2.8 µA 0.16 mA/A + 31 µA 0.29 mA/A + 31 µA 0.42 mA/A + 31 µA	Multifunction Calibrator
	(10 to 16.5) A (16.5 to 150) A (150 to 1 000) A	4.7 mA/A + 29 mA 4.7 mA/A + 0.21 A 4.7 mA/A + 0.99 A	Multifunction Calibrator with Current Coil
DC Current - Measure	Up to 100 nA 100 nA to 1 µA (1 to 10) µA (10 to 100) µA 100 µA to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	16 µA/A + 45 pA 11 µA/A + 54 pA 20 µA/A + 0.1 nA 20 µA/A + 0.81 nA 15 µA/A + 14 nA 20 µA/A + 51 nA 35 µA/A + 0.51 µA 35 µA/A + 5.5 µA/A	High Resolution DMM



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Current - Source	(29 to 330) μ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz 330 μ A to 3.3 mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (3.3 to 33) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (33 to 330) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz 330 mA to 3 A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (3 to 11) A (45 to 100) Hz 100 Hz to 1 kHz (5 to 10) kHz	1.6 mA/A + 78 nA 1.2 mA/A + 78 nA 0.97 mA/A + 78 nA 2.3 mA/A + 0.12 μ A 6.2 mA/A + 0.16 μ A 12 mA/A + 0.31 μ A 1.6 mA/A + 0.13 μ A 0.97 mA/A + 0.12 μ A 0.78 mA/A + 0.12 μ A 1.6 mA/A + 0.16 μ A 3.9 mA/A + 0.23 μ A 7.8 mA/A + 0.47 μ A 1.4 mA/A + 1.6 μ A 0.7 mA/A + 1.6 μ A 0.31 mA/A + 1.6 μ A 0.62 mA/A + 1.6 μ A 1.6 mA/A + 2.3 μ A 3.1 mA/A + 3.1 μ A 1.4 mA/A + 16 μ A 0.7 mA/A + 16 μ A 0.31 mA/A + 16 μ A 0.78 mA/A + 39 μ A 1.6 mA/A + 78 μ A 3.1 mA/A + 0.16 mA 1.4 mA/A + 78 μ A 0.47 mA/A + 78 μ A 4.7 mA/A + 0.78 mA 19 mA/A + 3.9 mA 0.45 A/A + 1.9 mA 0.77 A/A + 1.6 mA 23 mA/A + 1.6 mA	Multifunction Calibrator



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Current - Source	(10 to 16.5) A (45 to 65) Hz (65 to 440) Hz (16.5 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	5.5 mA/A + 33 mA 10 mA/A + 35 mA 5.6 mA/A + 0.27 A 10 mA/A + 0.27 A 5.1 mA/A + 1.7 A 12 mA/A + 1.1 A	Multifunction Calibrator with Current Coil
AC Current - Measure	100 µA to 1 mA (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz (10 to 100) mA (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz	1.5 mA/A + 0.2 µA 0.6 mA/A + 0.2 µA 0.3 mA/A + 0.2 µA 4 mA/A + 2 µA 1.5 mA/A + 2 µA 0.6 mA/A + 2 µA 0.3 mA/A + 2 µA 4 mA/A + 20 µA 1.5 mA/A + 20 µA 0.6 mA/A + 20 µA 0.3 mA/A + 20 µA	High Resolution DMM
AC Current - Measure	100 mA to 1 A (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz	4 mA/A + 0.2 mA 1.6 mA/A + 0.2 mA 0.6 mA/A + 0.2 mA 1 mA/A + 0.2 mA	High Resolution DMM
Resistors - Source Fixed Values (at 1 kΩ)	24.9 Ω 375.6 Ω 5.97 kΩ 95.3 kΩ	6.9 mΩ 51 mΩ 0.79 Ω 12 Ω	Standard Resistors Kit
Multifunction Calibrator	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ	27 µΩ/Ω + 1.2 mΩ 19 µΩ/Ω + 1.7 mΩ 20 µΩ/Ω + 1.4 mΩ 21 µΩ/Ω + 2 mΩ 22 µΩ/Ω + 1.8 mΩ 20 µΩ/Ω + 22 mΩ 22 µΩ/Ω + 17 mΩ 19 µΩ/Ω + 0.3 Ω	Multifunction Calibrator



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Resistance - Source	(33 to 110) kΩ (110 to 330) kΩ 330 kΩ to 1.1 MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	19 µΩ/Ω + 0.52 Ω 24 µΩ/Ω + 2.7 Ω 24 µΩ/Ω + 3.8 Ω 24 µΩ/Ω + 0.1 kΩ 95 µΩ/Ω + 0.12 kΩ 0.17 mΩ/Ω + 3.1 kΩ 0.38 mΩ/Ω + 3.7 kΩ 2.3 mΩ/Ω + 81 kΩ 12 mΩ/Ω + 0.4 MΩ	Multifunction Calibrator
Resistance - Measure	Up to 10 Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	14 µΩ/Ω + 75 µΩ 12 µΩ/Ω + 0.52 mΩ 10 µΩ/Ω + 0.57 mΩ 9.7 µΩ/Ω + 13 mΩ 10 µΩ/Ω + 57 mΩ 15 µΩ/Ω + 2.1 Ω 49 µΩ/Ω + 0.12 kΩ 0.16 mΩ/Ω + 79 kΩ 4.5 mΩ/Ω + 0.56 MΩ	High Resolution DMM
RTD Simulation	Pt 385 (100 Ω) (73 to 1 073) K (-200 to 800) °C Pt 385 (1 000 Ω) (73 to 903) K (-200 to 630) °C Pt 3916 (100 Ω) (73 to 903) K (-200 to 630) °C Pt 3926 (100 Ω) (73 to 903) K (-200 to 903) °C Ni 120 (120 Ω) (193 to 533) K (-80 to 260) °C Pt 385 (200 Ω) (73 to 903) K (-200 to 630) °C Pt 385 (500 Ω) (73 to 903) K (-200 to 630) °C	0.09 K (0.09 °C) 0.1 K (0.1 °C) 0.09 K (0.09 °C) 0.09 K (0.09 °C) 0.13 K (0.13 °C) 0.1 K (0.1 °C) 0.1 K (0.1 °C)	High Resolution DMM



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Capacitance – Source			
10 Hz to 1 kHz	(3.3 to 11) nF	1.8 mF/F + 11 pF	
Charge/Discharge Rate	(11 to 33) nF	1.7 mF/F + 95 pF	
10 Hz to 600 Hz	(33 to 110) nF	1.7 mF/F + 0.13 nF	
Charge/Discharge rate	(110 to 330) nF	1.4 mF/F + 0.57 nF	
10 Hz to 300 Hz	(0.33 to 1.1) μ F	1.9 mF/F + 0.91 mF	
Charge/Discharge rate	(1.1 to 3.3) μ F	1.4 mF/F + 5.7 nF	
10 Hz to 150 Hz	(3.3 to 11) μ F	1.8 mF/F + 10 nF	Multifunction Calibrator
Charge/Discharge rate	(11 to 33) μ F	2.6 mF/F + 52 nF	
10 Hz to 120 Hz	(33 to 110) μ F	3.4 mF/F + 88 nF	
Charge/Discharge rate	(110 to 330) μ F	3 mF/F + 0.51 μ F	
Up to 50 Hz	330 μ F to 1 mF	3.4 mF/F + 0.99 μ F	
Charge/Discharge rate			
Up to 50 Hz			
Charge/Discharge rate			
Capacitance - Source Fixed Values	1 nF 1 μ F 1 kHz 100 μ F	0.28 nF 1.5 nF 1.5 nF 1.5 nF 15 nF 15 nF 0.15 μ F 0.15 μ F 0.15 μ F	Standard Capacitors



ANSI-ASQ National Accreditation Board

ANAB

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Capacitance - Source Fixed Values	1 mF 10 mF	1.9 μ F 2.1 μ F 2.1 μ F 0.11 mF 0.11 mF 0.15 mF	Standard Capacitors
Inductance - Source Fixed Values	10 mH	6.5 μ H 6.5 μ H	Standard Inductors
Oscilloscopes Bandwidth (Leveled Sine Wave)	50 kHz to 600 MHz	(4.8 + 0.006 8X) %	
DC Voltage 50 Ω load 1 M Ω load	(0 to 6.6) V (0 to 130) V	1.9 mV/V + 0.37 mV 0.46 mV/V + 0.5 mV	Multifunction Calibrator
Square Wave -Amplitude 50 Ω load 1 M Ω load	(0 to 6.6) V (0 to 130) V	0.19 mV/V + 0.46 mV 0.77 mV/V + 0.64 mV	
Rise Time	3.5 ns Pulse Edge	41 ps	
Oscilloscopes Time Marker	(2 to 10) ns (20 to 100) ns (100 to 500) ns (1 to 20) ms (50 to 500) ms (1 to 5) s	2.9 ns/s + 7.8 ps 27 ns/s + 7.7 ps 0.15 μ s/s + 7.7 ps 4.6 ns/s + 8.6 ns 1.4 ns/s + 44 ns 2.8 ms/s + 9 ms	Multifunction Calibrator

Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
RF Absolute Power - Generate Sine Wave into 50 Ω 200 Hz to 81 MHz Sine Wave into 75 Ω 200 Hz to 81 MHz	(-86.98 to 13.01) dBm (-88.74 to 11.25) dBm	0.015 dBm 0.015 dBm	Synthesized Level Generator



ANSI-ASQ National Accreditation Board

Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
RF Absolute Power - Generate <1.0 Hz to 100 KHz (0.10 to 20) MHz	(-56 to 23) dBm	0.33 dBm	Function Generator
RF Absolute Power - Generate 100 kHz to 2.060 GHz	(-140 to 13) dBm	0.099 dBm	Signal Generator
RF Absolute Power - Generate 10 MHz to 2.3 GHz	(-9.95 to 10) dBm	0.018	Signal Generator.
RF Absolute Power - Generate (2.3 to 26.5) GHz	(-79.95 to -10) dBm	0.061 dBm	Signal Generator
RF Absolute Power - Generate (20 to <26.5) GHz	(-100 to -80) dBm	0.074 dB	Signal Generator
RF Absolute Power - Generate ≥ 10 MHz to ≤ 40 GHz	(-120 to 20) dBm	0.125 dBm	Signal Generator
RF Absolute Power Measure 10 MHz to 18 GHz 10 MHz to 18 GHz	(0 to 35) dBm (35 to 44) dBm	0.12 dBm 0.23 dBm	RF Power Meter and Power Sensor
0.1 MHz to 4.2 GHz 0.1 MHz to 4.2 GHz	(-30 to 10) dBm (10 to 20) dBm	0.08 dBm 0.09 dBm	
10 MHz to 18 GHz 10 MHz to 18 GHz	(-70 to -30) dBm (-30 to -20) dBm	0.1 dBm 0.11 dB	
10 MHz to 18 GHz 10 MHz to 18 GHz	(-30 to 10) dBm (10 to 20) dBm	0.08 dBm 0.17 dBm	
RF Absolute Power Measure 30 MHz to 26.5 GHz 30 MHz to 26.5 GHz	(-20 to -10) dBm (-10 to 30) dBm	0.15 dBm 0.15 dBm	RF Power Meter and Power Sensor

Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Micrometers – Outside ³	Up to 4 in (4 to 20) in (20 to 36) in	$(44 + 16L) \mu\text{in}$ $(32 + 22L) \mu\text{in}$ $(32 + 22L) \mu\text{in}$	Gage Blocks, Optical Flats



ANSI-ASQ National Accreditation Board

ANAB

Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Calipers - Outside Jaws ³	Up to 6 in (6 to 40) in	(420 + 3.9L) μ in (350 + 17L) μ in	Gage Blocks
Calipers - Inside Jaws ³	Up to 24 in (24 to 40) in	(480 + 3.9L) μ in (340 + 19L) μ in	Reference Bar, Gage Blocks
Calipers - Depth ³	Up to 24 in	(530 + 1.7L) μ in	Gage Blocks, Surface Plate
Height Gages ³	Up to 24 in (24 to 40) in	(490 + 10L) μ in (260 + 19L) μ in	Reference Bar, Surface Plate, Test Indicator
Micrometers - Inside ³ (Head Movement Only)	Up to 1 in	(81 + 24L) μ in	Gage Blocks, Gage Holder
Micrometers - Inside ³ (Resolution 0.0001 in) (Resolution 0.001 in)	Up to 6 in (6 to 24) in (24 to 40) in	(100 + 12L) μ in (38 + 22L) μ in (460 + 17L) μ in	Reference Bar, Gage Blocks
Micrometers - Depth ³	Up to 12 in	(630 + 4.5L) μ in	Gage Blocks, Surface Plate
Bore Gages ³ (Resolution 0.000 1 in)	(0.1 to 0.5) in (0.5 to 3) in	(80 + 2L) μ in (150 + 19L) μ in	Master Ring Gages
Indicators ³ Test, Dial, Digital (Resolution 0.000 1 in)	Up to 2 in	(68 + 25L) μ in	Gage Blocks, Calibration Tester, Surface Plate
Flatness	Up to 4 in	5.5 μ in	Master Flat
Optical Comparator ³ Horizontal Readout Vertical Readout	Up to 8 in Up to 8 in	(740 + 8.6L) μ in (760 + 8.7L) μ in	Reading Scale
Thickness (Feeler) Gages	(0 to 0.05) in	125 μ in	Digital Micrometer
Rulers ³	Up to 40 in	(3 200 + 112L) μ in	Caliper
Plain Plugs ²	Up to 90 mm	4.3 μ m + 0.002 8 μ m/mm	IAC Master Scanner
Plain Ring ²	2.5 to 100 mm	4.3 μ m + 0.004 7 μ m/mm	IAC Master Scanner
Thread Flank Angle Measurements ²	Up to 60 °	0.15 deg + 0.000 073 deg/deg	IAC Master Scanner



Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Thread Plug Gages ² Major / Minor Diameter	Up to 90 mm	4.2 μm + 0.016 $\mu\text{m}/\text{mm}$	IAC Master Scanner
Effective Pitch Diameter	Up to 90 mm	4.3 μm + 0.016 $\mu\text{m}/\text{mm}$	
Pitch	0.1 to 40 mm	1.4 μm + 0.003 7 $\mu\text{m}/\text{mm}$	
Thread Ring Gages ² Major / Minor Diameter	2.5 to 100 mm	4.9 μm + 0.012 $\mu\text{m}/\text{mm}$	IAC Master Scanner
Effective Pitch Diameter	2.5 to 100 mm	5.1 μm + 0.01 $\mu\text{m}/\text{mm}$	
Pitch	0.1 to 40 mm	1.6 μm + 0.003 3 $\mu\text{m}/\text{mm}$	

Mass

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Balances	Up to 410 g	0.016 mg/g + 1.6 mg	Class 3 Weights
	Up to 9 kg (20 lb)	0.092 mg/g + 0.1 g 0.000 092 lb/lb + 0.000 22 lb	Class 6 Weights
Scales	Up to 400 lb	0.24 lb + 0.000 21 lb/lb	Class 6 Weights
Torque Tools	(0.4 to 2) Nm (4 to 18) lbf in	0.000 56 Nm/Nm + 0.000 034 Nm 0.000 56 lbf in/lbf in + 0.003 lbf in	Torque Tester
Torque Tools	(2.26 to 11.29) Nm (20 to 100) lbf in (67 to 338.9) Nm (50 to 250) lbf ft (271.1 to 1 356) Nm (200 to 1 000) lbf ft	0.034 Nm + 0.002 6 Nm/Nm 0.3 lbf in + 0.002 3 lbf in / lbf in 0.22 Nm + 0.007 3 Nm/Nm 0.16 lbf ft + 0.005 36 lbf ft / lbf ft 0.46 Nm + 0.078 1 Nm/Nm 0.34 lbf ft + 0.005 76 lbf ft/lbf ft	Torque Transducer, Torque Display
Tensiometers ^{2,3}	(5 to 600) lbf	(1.6 + 0.034F) lbf	Class 6 Weights
Force Gage ²	Up to 1 000 lbf	2 lb + 0.000 46 lb/lb	Load cell with indicator
Pressure - Pneumatic	(-12 to 0) psi (0 to 30) psi (30 to 100) psi (100 to 1 000) psi	0.008 2 psi + 64 $\mu\text{psi}/\text{psi}$ 0.000 7 psi + 98 $\mu\text{psi}/\text{psi}$ 0.008 2 psi + 64 $\mu\text{psi}/\text{psi}$ 0.081 psi + 78 $\mu\text{psi}/\text{psi}$	Precision Pressure Controller used as Standard; Calibration Media -Nitrogen



ANSI-ASQ National Accreditation Board

Mass

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Pressure ² - Hydraulic, Cross Floating	41.4 kPa to 16.5 MPa (6 to 2 400) psi	0.22 kPa + 0.19 Pa/Pa 0.032 psi to 0.000 028 psi/psi	Comparison to Ruska 2400 Standard Dead Weight Tester
	207 kPa to 82.7 MPa (30 to 12 000) psi	0.35 kPa + 0.25 Pa/Pa 0.05 psi + 0.000 037 psi/psi	
	Up to 499.86 MPa (Up to 72 500) psi	4.8 MPa + 0.000 072 MPa/MPa 700 psi + 0.000 000 5 psi/psi	
Pressure - Hydraulic	Up to 499.86 MPa (Up to 72 500) psi	4.8 MPa + 0.000 072 MPa/MPa 700 psi + 0.000 000 5 psi/psi	Pressure Transducer
	(34.48 to 137.92) MPa (5 000 to 20 000) psi	77 kPa + 0.000 55 Pa/kPa 11.2 psi + 0.000 08 psi/psi	Precision Pressure Monitor
Hardness Testers	HRBW Scale		
	Low	2 HRBW	Indirect Comparison to Hardness Test Blocks
	Mid	1.6 HRBW	
	High	1.8 HRBW	
	HRC Scale		
	Low	1.3 HRC	
Rockwell Superficial	Mid	1.3 HRC	
	High	1.1 HRC	
	HR30TW Scale		
	Low	1.6 HR30TW	
	Mid	1.3 HR30TW	
	High	1.4 HR30TW	

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Temperature at Ice Point	273.15 K (0 °C)	0.033 K 0.033 °C	Standard Multimeter and Platinum Resistance Thermometer
Temperature - Measure	(73 to 933) K (-200 to 660) °C	0.000 004 K/K + 0.032 K 0.000 004 °C/°C + 0.032 °C	Standard Multimeter and Platinum Resistance Thermometer
	(0 to 1 750) °C	0.004 3 °C/°C + 0.57 °C	Type R Thermocouple and Multifunction Calibrator



ANSI-ASQ National Accreditation Board

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Humidity - Source	(10 to 95) %RH	0.81 %RH + 0.014 %RH/%RH	Humidity Chamber and Humidity Meter
Humidity - Measure	(10 to 95) %RH	0.7 %RH + 0.015 %RH/%RH	

Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Frequency - Source Using Calibrator's Normal Output	(0.01 to 120) Hz 120 Hz to 1.2 kHz (1.2 to 12) kHz (12 to 120) kHz 120 kHz to 1.2 MHz (1.2 to 2) MHz	1.2 μ Hz/Hz + 0.11 mHz 1.6 μ Hz/Hz + 0.48 mHz 1.9 μ Hz/Hz + 0.4 mHz 1.9 μ Hz/Hz + 1.2 mHz 1.9 μ Hz/Hz + 1.2 mHz 1.9 μ Hz/Hz + 14 mHz	Multifunction Calibrator
Frequency - Source Using Calibrator's Oscilloscope Output	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	1.9 μ Hz/Hz + 5.1 Hz 1.9 μ Hz/Hz + 0.3 Hz 1.9 μ Hz/Hz + 70 mHz	
Frequency - Measure	1 Hz to 10 MHz	0.5 mHz/Hz + 0.1 μ Hz	High Resolution DMM
Stopwatches	Up to 24 hours	0.19 s	NIST UTC Phone Time Signal
Frequency – Source and Measure	0.1 μ Hz to 3 GHz 10 Hz to 26.5 GHz	0.012 Hz + 0.385 Hz/MHz 0.17 Hz	Frequency Counter

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. Mobile and On-site calibration service is available for most parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope. Marsh Mobile Lab environmental conditions are controlled to meet the conditions needed to achieve the uncertainties listed in the Calibration and Measurement Capability (CMC) of the laboratory.
2. These parameters can only be performed at the laboratory's fixed location.
3. X = measured value, F = measured value in lbf, L = Length in inches.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1172.



R.D. Sauer

Vice President